



## VEEG Information Sheet

Patient Name: \_\_\_\_\_

Schedule Test Date: \_\_\_\_\_

Return Equipment Date: \_\_\_\_\_

**Edwin Liu, M.D.**

**Bernardo Flasterstein, M.D.**

**Farjam Farzam, M.D.**

**Abigail Ley, M.D.**

**Novette Green, A.R.N.P.**

**Michelle L. Jampol, A.R.N.P.**

**Alyssa P. Ausnehmer, A.R.N.P.**

\_\_\_\_\_  
**12959 Palms West Drive  
Suite 120  
Loxahatchee, FL 33470**

\_\_\_\_\_  
**11310 Legacy Avenue  
Palm Beach Gardens, FL  
33410**

\_\_\_\_\_  
**150 SW Chamber Court  
Suite 203  
Pt St Lucie, FL 34986**

\_\_\_\_\_  
**Phone 561-753-8888  
Facsimile 561-795-5004  
www.pedineurologists.com**

### Instructions:

- 1. Your child will be connected to a monitor for 24 hours. They will not be able to attend school and you may need to miss work to assure the test is complete.**
2. Although the camera is portable, keep camera plugged in as much as possible and in the view of the patient
3. If leads disconnect during the test, DO NOT attempt to reattach. If more than half of the leads are disconnected, discontinue exam
4. Once test has reached 24 hours, remove leads (instructions below)
5. Do NOT attempt to turn off or tamper with the equipment
6. Equipment MUST be returned when instructed and before 2:00pm
- 7. Fill out Event Log Sheet and return with equipment**

### Removing Leads:

Carefully peel off any tape on the forehead and gently pull off leads from head. Remove any pieces of gauze and patches left on leads. Place electrodes in a plastic bag or wrap tips in paper towel and place inside the book bag provided. Wash hair with warm water to allow paste to dissolve.

Reminder: Staff will not remove leads at the office. Please remember to return the camera charger with the equipment.

Technician's Emergency Contact: Carlos (786) 597-4990

A **\$300.00** per day fee will be accessed to your account if equipment is not returned by the specified date given by the technician. In addition, the patient is responsible for any loss or damage to the equipment.

By signing I stated I am fully aware of the information listed above:

Parent Signature: \_\_\_\_\_